

온도기록지

Temperature Log Sheet

2026년 7월

위치 (Location) : Core lab			장비명		Deepfreezer		
			모델명		Forma 900 series (908)		
			Serial No.		837903-515		
교정 번호(Calibration No.) : 2600075-0001 교정 일자(Calibration date.) : 2026.01.19							
날짜	요일	측정시각	온도 Temp °C			확인자 sign	비고
			Max.	Min.	Actual		
1	수	08:11 A.M. / P.M.	-75.5	-78.1	-76.6	h77	
2	목	08:15 A.M. / P.M.	-72.3	-78.0	-76.5	h77	
3	금	08:16 A.M. / P.M.	-75.1	-78.0	-76.1	h77	
4	토	A.M. / P.M.					
5	일	A.M. / P.M.					
6	월	08:16 A.M. / P.M.	-74.8	-78.3	-77.6	h77	
7	화	08:10 A.M. / P.M.	-74.5	-78.2	-78.0	h77	
8	수	08:10 A.M. / P.M.	-74.6	-78.2	-76.5	h77	
9	목	08:40 A.M. / P.M.	-75.3	-78.0	-76.4	h77	
10	금	08:20 A.M. / P.M.	-74.8	-77.9	-77.9	h77	
11	토	A.M. / P.M.					
12	일	A.M. / P.M.					
13	월	08:25 A.M. / P.M.	-75.2	-78.1	-77.9	h77	
14	화	08:25 A.M. / P.M.	-75.4	-78.0	-76.2	h77	
15	수	08:26 A.M. / P.M.	-74.1	-78.0	-75.9	h77	
16	목	08:25 A.M. / P.M.	-75.0	-77.9	-76.2	h77	
17	금	A.M. / P.M.				24시간 작동. 제1222	
18	토	A.M. / P.M.				2026.07.20 / 21.	
19	일	A.M. / P.M.					
20	월	08:25 A.M. / P.M.	-74.3	-78.1	-76.6	h77	
21	화	08:18 A.M. / P.M.	-74.9	-77.9	-77.6	h77	
22	수	08:30 A.M. / P.M.	-74.1	-78.3	-76.9	h77	
23	목	08:30 A.M. / P.M.	-74.0	-78.1	-76.5	h77	
24	금	08:30 A.M. / P.M.	-74.9	-78.0	-76.7	h77	
25	토	A.M. / P.M.					
26	일	A.M. / P.M.					
27	월	08:22 A.M. / P.M.	-75.6	-78.1	-75.7	h77	
28	화	08:26 A.M. / P.M.	-75.2	-78.1	-76.2	h77	
29	수	08:20 A.M. / P.M.	-75.4	-78.1	-77.7	h77	
30	목	08:20 A.M. / P.M.	-74.8	-78.1	-77.1	h77	
31	금	08:20 A.M. / P.M.	-75.4	-78.0	-76.6	h77	